

Air-Dryable™ qPCR Mix, 4x

For research or further manufacturing use only

Catalog No:	MDX082
Lot No:	B364130
Storage Conditions:	-20°C
Component Lot No:	526101A
Expiry date:	February 2028

Quality Control Parameters

Analysis	Specification	Result
Functional	<p>Amplification of a target from a manufactured DNA template using a SYBR Green based qPCR assay under standard cycling conditions.</p> <p><u>Pass Criteria:</u></p> <p>Amplification profiles must be consistent for the test and reference sample within ± 0.5 Cq difference.</p>	Passed
DNA contamination	<p>Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked. Test sample must amplify in concordance with control sample.</p> <p><u>Pass Criteria:</u></p> <p>Amplification traces must overlay with the negative control.</p>	Passed
DNase contamination	<p>DNase contamination is measured as DNA substrate degradation against a DNase I dilution series by agarose gel electrophoresis.</p> <p>Limit of detection: 6.25×10^{-4} KU DNase I.</p> <p><u>Pass Criteria:</u></p> <p>No detectable degradation.</p>	Passed

QA / QC Representative:



J. Rahnenführer

 Date: 23rd January 2026

United Kingdom

 Tel: +44 (0)20 8830 5300
 Fax: +44 (0)20 8452 2822

USA

 Tel: +1 901.382.8716
 Fax: +1 901.382.0027

Germany

 Tel: +49 (0)3371 60222 00
 Fax: +49 (0)3371 60222 01