

RNase Inhibitor (Glycerol Free)

Suitable for Research and further Manufacturing Use

Catalog No:	MDX120
Lot No:	B362030
Storage Conditions:	-80°C
Component Lot No:	326501A
Expiry date:	February 2028

Quality Control Parameters

Inhibits a broad spectrum of eukaryotic RNases, including RNases A, B and C to control for contaminants in RT-PCR assays.

Analysis	Specification	Result
Performance test	Test level of inhibition by incubating total RNA with concentration gradient of RNase A. Bands were observed with agarose gel electrophoresis (ethidium stained).	Passed
Activity	Activity is measured by inhibition of the hydrolysis of cytidine 2',3'-cyclic monophosphate (2',3' cCMP) by RNase A Pass Criteria: ≥ 40 U/ μ L	Passed
Specific Activity	≥ 80000 U/mg	Passed
Purity	Densitometric analysis of SDS-Page. Pass Criteria: $> 95\%$	Passed
Bioburden	< 50 CFU/mL	Passed
DNA Content	≤ 100 pg/mg of protein	Passed
Nicking Activity	Incubation of dNTP mix with supercoiled control plasmid. Analysed by agarose gel electrophoresis. Test sample does not show an increase of linearized or relaxed plasmid.	Passed
DNase contaminations	Incubation of a 1Kb double stranded DNA fragment. Incubation for 4 hours at 37°C with dilution series of DNase I. Analysed by agarose gel electrophoresis. Test sample must show less degradation than the limit of detection 2.5×10^{-3} U DNase.	Passed
RNase contamination	Quantitative PCR analysis with high and low RNase standards. Test sample must show less RNase activity than the limit of detection 9.7×10^{-3} ng/ μ L RNase.	Passed
RNase activity after heating	Quantitative PCR analysis with high and low RNase standards. Test sample must show less RNase activity than the limit of detection 9.7×10^{-3} ng/ μ L RNase.	Passed

 QA / QC Representative: 

X. Chen

 Date: 05th January 2026

United Kingdom

 Tel: +44 (0)20 8830 5300
 Fax: +44 (0)20 8452 2822

USA

 Tel: +1 901.382.8716
 Fax: +1 901.382.0027

Germany

 Tel: +49 (0)3371 60222 00
 Fax: +49 (0)3371 60222 01