


	Certificate of Analysis	COA No: CA_BSM-0119
		Version: 02

Lyo-Ready™ Direct RNA-DNA qPCR Blood, 4x For research or further manufacturing use only	Catalog No:	MDX123
	Lot No:	B356530
	Storage Conditions:	-20°C
	Component Lot No:	525210A
	Expiry date:	November 2027

Quality Control Parameters

Analysis	Specification	Result
Functional	<p>Quantitative real-time PCR analysis amplifying a target gene from dilutions of mouse RNA under standard cycling conditions.</p> <p><u>Pass Criteria:</u></p> <p>Amplification profile of a 1:10 dilution must be consistent for the test and reference sample within ≤ 0.5 Cq difference.</p> <p>The end florescence of the 1:10 dilution must be consistent for the test and reference sample within ≤ 0.10 difference.</p>	Passed
DNA contamination	<p>Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked. Test sample must amplify in concordance with control sample.</p> <p><u>Pass Criteria:</u></p> <p>Amplification traces must overlay with the negative control.</p>	Passed
DNase contamination	<p>Incubation of a 1Kb double stranded DNA fragment. Incubation for 4 hours at 37°C with dilution series of DNase I. Analysed by agarose gel electrophoresis. Test sample must show less degradation than the limit of detection 2.5×10^{-3} U DNase.</p> <p><u>Pass Criteria:</u></p> <p>No detectable degradation.</p>	Passed
RNase contamination	<p>Quantitative PCR analysis with high and low RNase standards.</p> <p>Limit of detection: 9.7×10^{-3} ng/μL RNase</p> <p><u>Pass Criteria:</u></p> <p>Test sample must show less RNase activity than the limit of detection.</p>	Passed

QA / QC Representative: 

X.Chen

Date: 28th October 2025

United Kingdom

Tel: +44 (0)20 8830 5300
Fax: +44 (0)20 8452 2822

USA

Tel: +1 901.382.8716
Fax: +1 901.382.0027

Germany

Tel: +49 (0)3371 60222 00
Fax: +49 (0)3371 60222 01